

# **Element Ankle Brace**

Single Patient Use Only • Rx Only

#### **IMPORTANT INFORMATION**


Please read all instructions, warnings, and precautions before use. Correct application is essential for proper functioning of product.

# **INTENDED USE**


Acute ankle sprains, chronic ankle instabilities, high ankle sprains, or prophylactic use.

# PRECAUTIONS

This product is to be fitted by a physician (or properly licensed practitioner) who is familiar with the purpose for which they are responsible. The physician or practitioner is responsible for providing wearing instructions and precautions to other healthcare practitioners or healthcare providers involved in the patient's care and the patient. Consult your physician (or properly licensed practitioner) immediately if you experience sensation changes, unusual reactions, swelling or increased pain while using this product. Follow the instructions of your physician (or properly licensed practitioner) for length and duration of use.



# INSTRUCTIONS FOR USE



Before application, place an absorbent sock or ankle wrap on patient's ankle. Make sure the heel is as far back in the brace as possible. The criss-cross straps inside the brace located at the back of the foot should cup the heel.

 Secure the upper then lower calf strap for a snug fit.
NOTE: These straps are to be secured in

be secured in opposite directions. Top straps clockwise (a). Bottom strap counter clockwise (b).

3. Next, secure the criss cross straps to the

uprights. Begin by firmly securing the strap on the arch side of the foot over the front of the ankle in an upward direction (a). Next, secure the strap on the





# **Element Ankle Brace**

Single Patient Use Only • **Rx Only** 

#### **IMPORTANT INFORMATION**


Please read all instructions, warnings, and precautions before use. Correct application is essential for proper functioning of product.

# INTENDED USE

Acute ankle sprains, chronic ankle instabilities, high ankle sprains, or prophylactic use.


# PRECAUTIONS

This product is to be fitted by a physician (or properly licensed practitioner) who is familiar with the purpose for which they are responsible. The physician or practitioner is responsible for providing wearing instructions and precautions to other healthcare practitioners or healthcare providers involved in the patient's care and the patient. Consult your physician (or properly licensed practitioner) immediately if you experience sensation changes, unusual reactions, swelling or increased pain while using this product. Follow the instructions of your physician (or properly licensed practitioner) for length and duration of use.




# **INSTRUCTIONS FOR USE**

1. Open all straps and spread the uprights apart to position the foot in the brace at a 90° angle. NOTE:



Before application, place an absorbent sock or ankle wrap on patient's ankle. Make sure the heel is as far back in the brace as possible. The criss-cross straps inside the brace located at the back of the foot should cup the heel.

 Secure the upper then lower calf strap for a snug fit. NOTE: These straps are to be secured in



opposite directions. Top straps clockwise (a). Bottom strap counter clockwise (b).

 $\ensuremath{\textbf{3.}}$  Next, secure the criss cross straps to the

uprights. Begin by firmly securing the strap on the arch side of the foot over the front of the ankle in an upward direction (a). Next, secure the strap on the





DeRoyal®

outside of the foot over the front of the ankle again pulling in an upward direction (b). **NOTE:** It is important to pull tightly in an upward direction maintaining tension on the strap until it is secured on the upright. The straps should form a

criss-cross configuration and be positioned flat against the front of the leg.

**4.** Check for adequate circulation. Readjust straps as needed for a secure and comfortable fit.



CARE INSTRUCTIONS

Wipe down brace with damp cloth and air dry.



Part #0-1826 ©2007 DeRoyal. All Rights Reserved. Revised 11/07 outside of the foot over the front of the ankle again pulling in an upward direction (b). **NOTE:** It is important to pull tightly in an upward direction maintaining tension on the strap until it is secured

on the upright. The straps should form a criss-cross configuration and

be positioned flat against the front of the leg. 4. Check for adequate

circulation. Readjust straps as needed for a secure and comfortable fit.

# **CARE INSTRUCTIONS**

Wipe down brace with damp cloth and air dry.



DeRoyal®

Manufactured by DeRoyal 200 DeBusk Lane Powell, TN 37849 U.S.A. 888.938.7828 865.938.7828

Part #0-1826 ©2007 DeRoyal. All Rights Reserved. Revised 11/07